



श्यामा प्रसाद मुखर्जी पोर्ट, कोलकाता
SYAMA PRASAD MOOKERJEE PORT, KOLKATA
Formerly Kolkata Port Trust

OFFICE OF THE CHIEF MEDICAL OFFICER
CENTENARY HOSPITAL, MEDICAL DEPARTMENT

Sub : Notice Inviting Tender for Supply, Delivery, Installation & Commissioning of 1(One) No. Mobile Radiography System (100mA) Under BUY BACK at Centenary Hospital of Syama Prasad Mookerjee Port Kolkata. Tender No. MED / PR / ADVT / 358 / 22-23/ 2446
श्यामा प्रसाद मुखर्जी पोर्ट कोलकाता के शताब्दी अस्पताल में बाय बैक के तहत 1 (एक) नंबर मोबाइल रेडियोग्राफी सिस्टम (100mA) की आपूर्ति, वितरण, स्थापना और कमीशनिंग के लिए निविदा आमंत्रित करने की सूचना।

Corrigendum

In terms of the Prebid meeting held on 06/3/2023 following has changes have been made in the NIT.

1. This Corrigendum shall be read in conjunction with the Bidding Document (including Notice Inviting Tender).
2. All other terms and conditions of the Bidding Document shall remain unchanged. Details of Corrigendum are given below :-

Technical Specification Appendix - B has been amended :

Existing	Amended
<p>Power Supply The unit should run on Single phase 230 Volts, 20 Amps, 50 Hz. Auto compensation of Line Voltage Compensation of +- 10%</p>	<p>Power Supply Offered unit should be operable on , Single phase , 230 Volt AC, 50HZ ,15 Ampere wall socket and line regulation $\pm 10\%$ with automatic compensation.</p>

13/3/2023

Devi Chap Yun
Dy. Chief Medical Officer
For Chef Medical Officer